

Manual Supplement

Manual Title: T5-600/T5-1000 I.C.
Part Number: 1631780
Print Date: Dec. 2001
Revision/Date: 3, 1/13
Supplement Issue: 2
Issue Date: 10/20
Page Count: 2

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 489

Under **Continuity & Resistance**, change:

From: Greater than
400 Ω

To: >40 Ω

Change #2, TP-122

Within the **Specifications** table, in the **Enclosure Rating** row, change the IP spec:

From: IP 52

To: IP 40